ELECTROMAGNETIC FIELD TESTER

EMF METER

Model: EMF-823 ISO-9001, CE, IEC1010

www.yalab.com.tw

02-2389-0101



20 micro Tesla200 micro Tesla2,000 micro Tesla

200 milli Gauss2,000 milli Gauss20,000 milli Gauss





The Art of Measurement

ELECTROMAGNETIC FIELD TESTER

EMF METER

Model: EMF-823 <u>www.YaLAB.com.tw</u> 02-2389-0101

FEATURE

- * The EMF tester is designed to provide user a quick, reliable and easy way to measure electromagnetic field radiation levels around power lines, home appliances and industrial devices.
- * Wide measuring ranges, 20/200/2,000 micro Tesla, 200/2,000/20,000 milli Gauss.
- * The EMF tester is a cost effective, hand-held instrument designed and calibrated to measure electromagnetic field radiation at different bandwidths down to 50 Hz/60 Hz.

APPLICATIONS

This EMF tester is specifically designed to determine the magnitude of electromagnetic field radiation generated by power lines, computer's monitor, TV sets, video machinery and many other similar devices.

SPECIFICATION

13 mm (0.5") LCD, 3 1/2 digits. Max. indication 199.9.
20 micro Tesla x 0.01 micro Tesla 200 micro Tesla x 0.1 micro Tesla 2,000 micro
Tesla x 1 micro Tesla
* 1 micro Tesla = 10 milli-Gauss
200 milli Gauss x 0.1 milli Gauss 2,000 milli Gauss x 1 milli Gauss
20,000 milli Gauss x 10 milli Gauss
30 Hz to 300 Hz.
Single axis.
± (4 % + 3 d) - 20 micro Tesla/200 milli Gauss range.
± (5 % + 3 d) - 200 micro Tesla/2,000 milli Gauss range.
± (10 % + 5 d) - 2,000 micro Tesla/20,000 milli Gauss range.
* Spec. accuracy tested under 50 Hz or 60 Hz.
Display shows " 1 " .
Approx. 0.4 second.
DC 9 V battery (006P, 6F22).
0 to 50 $^{\circ}\mathbb{C}$ (32 to 122 $^{\circ}\mathbb{F}$).
Less than 80% RH.
215 g/0.48 LB (including battery).
H.W.D 163 x 68 x 24 mm (6.4 x 2.7 x 0.9 inch).
Operation Manual 1 PC.

CAUTION OF ELECTROMAGNETIC FIELD EXPOSURE

- ✓ Claims by some scientists that long term exposure to electromagnetic field may be the cause of childhood leukemia & other forms of cancer.
- Complete answers to any of these and related questions are not currently available. At the present time the most common practice is to avoid excess exposure over long period of time.
- ✓ "Prudent Avoidance" as stated by the Environmental Protection Agency(EPA) U. S. A. is recommended.

EMF-823+0503

^{*} Appearance and specifications listed in this brochure are subject to change without notice.